

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 22 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130174
Name		Gary Lowe	ID# 7219 Date July 17, 2009
A	Agency	Haines PD	Phone # 766-2121
Instrument Location		Haines PD	
Alco S/N		X124631	Target Value .082 High Pressure 800
Alco Test Values		.082 1 st Alco	.082 2 nd Alco
Signature		Gary Lowe	CB3 7/24/09
(OVER)			


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of July 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130174

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JULY 17, 2009

OPERATOR'S NAME:

LOWE, GARY, A

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 7172009

DEPT/AGENCY: HNS1

CASE/REPORT: 09-71709

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X124631

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.24 in
ALCO TARGET .002 14:37
BLANK TEST .000 14:38
INTERNAL STANDARD VERIFIED 14:38
ALCO TV 30.24 in .002 14:38
BLANK TEST .000 14:39
SUBJECT SAMPLE .000 14:39
BLANK TEST .000 14:40
ALCO TV 30.24 in .002 14:40
BLANK TEST .000 14:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JULY 17, 2009

TIME 14:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.24 in
FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~■